

International University of Africa

Faculty of pure and applied sciences

**Design a Biological Phantom for Human Breast
(mammographic)**

Thesis submitted in partial fulfillment of academic requirement for the
degree of the master's

Presented by:

Naglaa Ahmed Abdalla

Supervised by:

Dr. Fawzia Elsadig Mohammed Elbashir

الإستهلال

قال الله تعالى:

(اقرأ بأسم ربك الذي خلق (1) خلق الإنسان
من علق (2) اقرأ وربك الأكرم (3) الذي
علم بالقلم (4) علم الإنسان ما لم يعلم
(5)).

سورة العلق: الآية (1-5)

Dedication

*I would like to dedicate this work to my
family and friends*

Acknowledgement

I am greatly indebted to my supervisor Dr. Fawzia Elsadig for her all positive guidance , help and encouragement that help me to make me dreams a reality.

Without her help and support this thesis will never see the light in this shape .

Moreover working with her has proven invaluable for focusing on how the planning process can be shaped and the reasons why it should be approached in certain ways.

I would also like to thanks my supervisor for her generosity in sharing her excellent insights in to the role of trust in resources management with me since I have started my research .

My thanks also goes to a number of colleagues who provided me with several of the ideas used in this master degree thesis .

I would like to extend my thanks to my colleagues at university for their extreme help, support, stimulations and encouragement in order to overcome all difficulties.

Finally I would like to thanks me family ,mainly mother and my late father , for their love , support and being patient to keep all my demand at my hand.

.

ABSTRACT

The goal of mammography is the detection, characterization, and evaluation of findings suggestive of breast cancer and other breast diseases. Annual screening mammography of age-appropriate asymptomatic women is currently the only imaging modality that has been proven to significantly reduce breast cancer mortality. A screening mammogram is an X-ray examination of the breast of an asymptomatic woman ,to perform QC charads . Exposure to high doses could be reason for suffering rather than diagnoses. Therefore, design biological phantom with similar properties to the actual human breast. This phantom helps radiology professional to ensure that their mammographic system is producing highest quality images.

The main objective of this research, to develop a new quality-control biological phantom composed of natural materials useful for mammography imaging modality. Two different locally available materials were tested using (GC-MS , Densitometer , mammography machine).The proposed phantom in this study was designed using egg white , egg shell , Beeswax and Animal fat to mimic the breast fat layer .Holder was designed with dimensions 10cm,10 cm and 4 cm from transparent plastic material.

Two Phantoms from materials that are available and cheap were Compared with APG Phantom .Good results of optical density average1.231 kg/m³was obtained from phantom one agree with results from patients of age less than 40 years. Phantom tow also obtained acceptable results average 1.141kg/m³in compared with a phantom APG from a Literature Review and also a great agreement with results from patients of age greater more than 40 years.

The study recommends future studies for mammography dosimetry and quality control measurements by using these phantoms.

المستخلص

الهدف من التصوير الاشعاعي للثدي هو الكشف ,وتوصيف وتقييم النتائج التي تدل علي سرطان الثدي وغيره من امراض الثدي.سنوياً التصوير الاشعاعي للثدي يحصر فئات عمرية من النساء ذات الامراض عديمه الاعراض وهو حالياً طريقة التصوير الوحيدة التي ثبتت لخفض كبير في معدل الوفيات من سرطان الثدي,يتم ذلك عبر ضبط الجودة . التعرض لجرعات عاليه يمكن ان يكون سبب معاناة بدلاً من التشخيص.

هذا الفانتوم الحيوي ذو خصائص مماثلة لسرطان الثدي الطبيعي , هذا الفانتوم يساعد الاشعة المهنية لضمان ان النظام بالاشعة ينتج الصورة في اعلي مستوي جودة.

وكان الهدف الرئيسي من هذا البحث هو تطوير تصميم مجسم حيوي جديد لمراقبة جودة أجهزة تصوير الثدي بالاشعة محلياً مما يساهم في الجانب الاقتصادي ويقلل تكلفة شرائه من الخارج .ولعمل ذلك قمنا بأختبار مجموعة من المواد بعدد من أجهزة القياس بأستخدام جهاز تصوير الثدي بالأشعة وكاشف لقياس الكثافة وجهاز تحليل المركبات, المواد الطبيعية المستخدمة (بياض وقشر البيض , وشمع النحل ,والشحوم الحيوانية) وذلك لتقليد المواد الاساسية المكونة للثدي الطبيعي , كما تم وضع هذه المواد في حاوية بلاستيكية شفافة .

وقد كانت نتائج هذا البحث ناجحه , حيث تم تصميم مجسمين بمواد طبيعيه ورخيصة ومتوفرة محلياً, وقد كانت نتائج المجسم الاول مقبولة حيث اعطي متوسط كثافة ضوئية 1.231 كيلوجرام\متر³ مقارنة مع صور المرضى ذو الاعمار الاقل من 40 سنة .

وعند أختبار المجسم الثاني, أعطي متوسط كثافة ضوئية 1.141 كيلوجرام\متر³ وهي اكثر دقه وجودة من المجسم الاول مقارنة بالمجسم المعياري (APG) والمرضى ذو الاعمار 40 سنه فما فوق .

نوصي بأستخدام هذا الفانتوم في قياس الجرعة الاشعاعية وقياس جودة الصورة QC.

List of Contents

Contents	Page
الاستهلال	I
DEDICATION	II
ACKNOWLEDGEMENTS	III
ABSTRACT	IV
المستخلص	V
LIST OF CONTENTS	VI
LIST OF FIGURES	IX
LIST OF TABLES	XI
List O F Abbreviation	XII
Chapter One: General Introduction	
1.1Introduction	1
1.2Problem Statement	3
1.3Objectives	3
1.3.1 General objective	3
1.3.2 Specific Objectives	3
1.4 Dissertation Layout	4

Chapter Two: Theoretical Background	
2.1 Anatomy of the breast	5
2.2Breast Materials	6
2.3Breast tissue simulation	7
2.4 Mammography Machine	8
2.4.1 Types of mammogram	11
2.5 Quality control (QC) in Mammography	12
2.6 Image Quality	13
2.6.1.Contrast	14
2.6.2.Noise	14
2.6.3. Signal-to-noise ratio (SNR)	14
2.Literature Review	14
Chapter Three :Material And Method	
3.1 Materials	18
3.2 DEVICE	20
3.2.1 Centrifuge Purchasing Guide	20
3.2.2 Photoelectric ColorimeterAP-101	21
3.2.3 Densitometer(341C)	21
3.2.4 Gas chromatography–mass spectrometry(GC MS)	23

3.2.5 Mammography	24
3.2.5.1 Main parts of machine	24
3.3 Methodology	25
3.3.1 The first test using the white material of the egg as protein material sample	25
3.3.2 The second test Beeswax	28
3.3.3 The third test Animal fat	29
3.4 Design of phantom	30
3.5 Holder design	33
Chapter Four :Results And Discussion	
4.1 Results	34
4.2 Discussion	38
Chapter Fife :Conclusion And Recommendations	
5.1 Conclusions	39
5.1 Recommendations	39
REFERENCES	40

LIST OF FIGURES

FIGURES	Page
Figure (2.1): The breast tissue anatomy .	6
Figure (2.2):The mammography machine .	9
Figure (2.3) : The x-ray device and compression plates used for mammograms.	10
Figure (2.4) :The breast mammogram.	10
Figure(2.5) :Digital VS film screen mammography	12
Figure (3.1):Egg white and eggshell	18
Figure (3.2):Beeswax	19
Figure (3.3):Animal fats	19
Figure (3.4): Centrifuge Purchasing Guide	20
Figure (3.5) :Photoelectric Colorimeter AP-101	21
Figure (3.6) :Portable Transmission Densitometer(341C)	22
Figure (3.7): Gas chromatography–mass spectrometry (GC-MS)	23
Figure (3.9): (1) The study sample ,(2)Protein STD,(3)Bivnet Reagut.	27
Figure(3.10): The time needed to separate the materials Beeswax	29
Figure(3.11): The time needed to separate the materials of animal fat	31
Figure (3.12): Phantom One	32

Figure (3.13):Phantom One in the mammography	32
Figure (3.14):Phantom Two	33
Figure (3.15):Phantom Two in the mammography	33
Figure (4.1): Phantom One mammography	35
Figure (4.2): Phantom Two mammography	36
Figure (4.3): Phantom(APG)in mammography	36

LIST OF TABLES

TABLES	Page
Table (2.1) : Shows fatty acid composition of human breast fat.	7
Table (3.1): Method of measuring protein in the study sample	26
Table(3.2): Analysis of Beeswax component	28
Table(3.3):Result of this preparation for the materials in the animal fat	29
Table (4.1): Compare the results between the materials used and the main components of human breast	37
Table (4.2): Density of Human Tissue study sample and phantom (APG) Measured by Densitometer.	38

List O F Abbreviation

SEM	Screen Film Mammography
NCI	National Cancer Institute
ACS	American Cancer Society
APG	Applied Physics Group
GC-MS	Gas Chromatography –Mass Spectrometry
TDLUS	Terminal Ductal lobular Units
MRI	Magnetic Resonance Imaging
AEC	Automatic Exposure Control
QC	Quality Control
ACR	American Collage of Radiology
BRTES	Breast Tissue Equivalent Series
MQSA	Mammography Quality Standards act
TLD	Thermo luminescent Dosimeter
KERMA (Ki)	Kinetic Energy Released Per Unit Mass
DBT	Digital Breast Tom synthesis
AGD	Average Glandular Breast Dose
GD	Glandular Dose

PVAL	Polyvinyl Alcohol
RCF	Relative Centrifugal Force
Gy	Gray
NCAT	4D NURBS-based Cardiac-Torso
GATE	Geant4 Application for Tomographic Emission
PMMA	Polymethyl methacrylate